


Squiggles TT Instructions – Engaging the Dynamic Base

To engage the dynamic base:


Turn the dials A on the Left and Right Hand Side of the rocker base so that the 'rocker' symbol B

aligns with the marker C.




While supporting the weight of the rocker bar, unclip the white buckle ${\color{red} \mathtt{D}}$, and allow the rocker bar

to drop to the ground.



Place your foot on the back of the rocker bar **E** and push downwards, meanwhile lift the Squiggles TT handles upwards and forwards **F**, until you hear a click. This is now engaged (blue on indicator).

Check: When the rocker is in the engaged position, the indicators G on the Left and Right Hand Side are blue.




Email: enquiries@sunrisemedical.com.au Website: www.sunrisemedical.com.au




Squiggles TT Instructions – Engaging the Dynamic Base

To disengage the dynamic base:

Turn the dials A on the Left and Right Hand Side of the rocker base so that the 'wheels on the ground' symbol B aligns with the marker C.



Place your foot on the back of the rocker bar D and push downwards, meanwhile lift the Squiggles TT handles upwards and forwards E, until you hear a click. Support the weight of the stander downwards until the wheels are on the ground.



Lift the rocker bar upwards and clip the white buckle F around the rocker bar.

Check: When the rocker is in the disengaged position, the indicators G on the Left and Right Hand

Side are white.